

1151 Boston-Providence Turnpike
 P.O. Box 9102 Norwood, MA 02062 USA
 T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

VF03 4abcdefghij2lm. Reflex Radar Level Meter BM 102
SF03 9abcdefghij2lm. Reflex Radar Level Meter BM 102
 IS / I, II, III / 1 / A, B, C, D, E, F, G / T* - F0820860412 / B; Entity
 I / 0 / AEx ia / IIC / T* - F0820860412 / B; Entity
 Entity Parameters
 Vmax = 30 V Imax = 150 mA Pmax = 1 W Ci = 10.5 nF Li = 10 µH

- a = probe type; 1, 2, 3, 4, 5, 6, 7
- b = probe material; 1, 2, 3, 4, 5, 6, 7, 8
- c = process connection
- d = probe length (meters)
- e = probe length (100mm)
- f = process pressure; 0, 1, 2
- g = sealing
- h = counter weight
- i = electrical connection; 1, 2, 3
- j = application; 1, 2
- l = precision
- m = flange temperature; 1, 2, 3,4

T*	Tambient	Option m
T6	60°C	-
T5	60°C	1
T4	60°C	1, 2
T3	55°C	1, 2, 3, 4
T3	60°C	1, 2, 3, 4

Equipment Ratings:

Intrinsically Safe apparatus for use in Class I, II and III, Division 1, Groups A, B, C, D, E, F and G and Class I Zone 0 Group IIC hazardous (classified) locations in accordance with Entity requirements and Control Drawing F0820860412.

Approved for:

Krohne S.A.
Usine des Ors
26103 Roamsn Cedex
France


This certifies that the equipment described has been found to comply with the following Factory Mutual Research Approval Standards and other documents:

Class 3600	1998
Class 3610	1999
Class 3810	1989

Original Approval Job Identification: 3009543 Approval Granted: February 12, 2001

Subsequent Revision Reports / Date Approval Amended

Factory Mutual Research Corporation



W Calder
Assistant Vice President
Approvals Division

15 Feb 01
Date