



DECLARATION OF CONFORMITY

KROHNE Altometer, Kerkeplaat 12, 3313 LC Dordrecht, The Netherlands



We hereby declare under our sole responsibility as manufacturer that the following products are in compliance with the UK statutory requirements:

| | |
|---------------------------|-------------------------|
| Electromagnetic Flowmeter | WATERFLUX 3070 C |
|---------------------------|-------------------------|

The product safety regulations are fulfilled in accordance with listed designated standards or technical specifications as applicable:

| Regulation | | Designated standards |
|--------------|---|-------------------------------------|
| 2016 No.1091 | Electromagnetic Compatibility | EN 61326-1:2013 (EN 61326-2-3:2013) |
| 2012 No.3032 | The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment | EN IEC 63000:2018 |

According to the applicable regulations, the equipment marking contains or references the following information:

| Marking | | | | | |
|-----------------------|-----------------------|-------------|----------|---------------|--|
| Regulation Year & No. | Conformity Assessment | Certificate | Category | Approved Body | Marking |
| 2016 No.1091 | - | n.a. | | n.a. |  n.a. |
| 2012 No.3032 | Module A | n.a. | | n.a. |  n.a. |

The Netherlands, 10/2022



A.H. Boer
General Manager