

ATTESTATION OF CONFORMITY

Issued to: KROHNE Altometer
P.O. Box 110
3300 AC Dordrecht
The Netherlands

For the product: Battery powered electromagnetic flow meter

Trademark: KROHNE Altometer

Type: WATERFLUX
- Field (remote) version: WATERFLUX 3070F – IP68
(WATERFLUX 3000 sensor + IFC 070 signal converter)

Ratings: IP68
IPX8, 10 meter depth, 1500 hours

Manufactured by: Krohne Altometer
Kerkeplaat 12
3313 LC Dordrecht
The Netherlands

Subject: Degree of protection provided by enclosure

Requirements: IEC 60529:1989 + A1:1999 + A2:2013
EN 60529:1991 + A1:2000 + A2:2013

This Attestation is granted on account of an examination by DEKRA, the results of which are laid down in test report no 2189640.02-INC issued on 22 April 2016.

The examination has been carried out on one single specimen of the product, submitted by the manufacturer. The Attestation does not include an assessment of the manufacturer's production. Conformity of his production with the specimen tested by DEKRA is not the responsibility of DEKRA.

Arnhem, 22 April 2016

Number: 2189640.02A

DEKRA Certification B.V.

F.S. Strikwerda
Certification Manager

© Integral publication of this attestation and adjoining reports is allowed